

# REQUEST FOR QUOTATION FOR THE SUPPLY MEDICAL DIAGNOSTIC (SLIT LAMP, AND DIGITAL FUNDUS CAMERA)



## Jhpiego

Jhpiego creates and delivers transformative health care solutions that save lives. In partnership with national governments, health experts and local communities, Jhpiego builds health providers' skills and develops systems that save lives now and guarantee healthier futures for women and their families.

### Close

#### Sector

Health

#### Rate this employer

Average: 4.2 (25 votes)

<http://www.jhpiego.org>

## REQUEST FOR QUOTATION

Reference Number: TN 002/Jhpiego-RW/FY25

### INTRODUCTION

#### Jhpiego Mission

Jhpiego enhances the health and saves the lives of women and families in limited-resource settings. For nearly four decades, we have put evidence-based health innovations into everyday practice to overcome barriers to high-quality health care services for the world's most vulnerable populations. From our origins as technical experts in reproductive, maternal and child health, Jhpiego has grown to embrace new challenges, including HIV/AIDS, malaria, GHS and cervical cancer prevention, reflecting the increasing interconnectedness of global health.

#### Jhpiego Values

At Jhpiego, we value our customers who have our respect, responsiveness, and commitment to excellence; our staff and global network of colleagues who bring cultural diversity, innovation, and a wealth of world experience; and our work culture, which is reflected in our team spirit, transparent communication, mutual respect, flexibility, and dedication.

Jhpiego hereby invites national bidders to submit quotations for the supply of the following **medical diagnostic (Slit Lamp, and Digital Fundus Camera)**. Participation in this

competition is open on equal conditions to all companies specialized in medical equipment supply and support.

## ITEMS, TECHNICAL SPECIFICATION, AND QUANTITIES

ITEM	TECHNICAL SPECIFICATIONS	QUANTITY
<b>Slit Lamp</b>	<ul style="list-style-type: none"> <li>• Magnification: 6x, 10x, 16x, 25x</li> <li>• Light Source: Xenon or LED</li> <li>• Illumination Types: Diffuse, direct, retroillumination, and slit illumination</li> <li>• Slit Width: 0.1 to 20 mm</li> <li>• Slit Length: Up to 15 mm</li> <li>• Slit Angle: Adjustable</li> <li>• Field of View: 12 mm x 16 mm</li> <li>• Filters: Red-free, blue cobalt, and UV filters</li> </ul>	<b>2</b>
<b>Digital Fundus Camera</b>	<p><b>Product &amp; Manufacturer Quality Standards:</b></p> <ul style="list-style-type: none"> <li>• The quoted model should be “USFDA approved (510k / CFG)” OR “European CE certified (where EU-CE certificate should be issued from Notified body having notified body number)” OR “BIS certified conforming to guideline specifically for the ‘Fundus Camera</li> <li>• The manufacturer of the quoted product should have “EN ISO 13485 certificate issued from a notified body</li> <li>• The quoted model should confirm to “IEC 60601” OR “IS/ ISO / IEC 80601 (Part 2)” OR “IS 13450 (Part 1)”.</li> <li>• All the Certificate/ISO to be provided</li> </ul> <p><b>Operational Requirement:</b> A fundus camera or retinal camera is a specialized low power microscope with an attached camera designed to photograph the interior surface of the eye, including the retina, optic disc, macula, and posterior pole (i.e. the fundus) A digital camera converts these images into digital images.</p> <p><b>Should meet the below technical Specification:</b></p> <ul style="list-style-type: none"> <li>• Field angles: 45 to 50 DEGREES or wider range</li> <li>• Image mask: minimum 45° to 50°, black, dark, transparent,</li> </ul>	<b>1</b>

- bright
- Capture modes: Color, green, Cobalt /blue, red filters
  - FFA and Red free, Cobalt or blue, Fundus Auto Fluorescence
  - Retinal Observation: Mydriatic : optical viewfinder, Non – Mydriatic : camera unit monitor
  - Angle of View - Mydriatic 50 degrees, Non-Mydriatic 45 degrees. Magnification - 2X (Digital)
  - Compensation of ametropia - +35 to – 35 D continuous
  - Minimum pupil size - Mydriatic Ø 4.5mm (approx), Non-Mydriatic Ø 3.5mm (approx.)
  - Working distance – front lens to patient's eye minimum 35mm
  - Camera Sensor - 15 Megapixels or higher
  - Range of base movement Minimum 60mm front and back, 100 mm side to side, 30mm up and Down (approx.)
  - Capture 1 chip sensor color
  - Monitor minimum 21" or more medical grade Full HD Monitor (1920 x 1080 or better resolution) for direct display.
  - Fixation Light –Internal and External fixation light both
  - Exposure interval 0.5 to 1 sec
  - Flash energy – Xenon flash lamp, 9 or more flash level
  - Frequency: 50 Hz
  - Facility for Data storage, data transfer, image archiving, image analysis
  - Instrument table Asymmetrical motorized suitable for patients in wheelchair.
  - PC (Minimum Specs: Core i5, 8GB RAM, 1TB Hard disk, 21 inch and above (1920 x 1080 or better resolution) medical grade LCD Full HD monitor)
  - Having highest number of programs in a machine with color, RF, FA, FAF, Blue, Red. P
  - Should be supplied with Machine Cover
  - The system should be with high quality and all accessories including table stand and UPS

**Power supply:**

- Should work with 220 to 240 V AC, 50Hz power supply.
- Should be provided with adapter with power cord of appropriate length having Rwanda plug.
- Should be supplied with over current circuit breaker to avoid high current surge.

#### **Management**

1. **Warranty:** Should have 02 years of manufacturer warranty excluding consumable parts.
2. The bidder shall provide a colored catalogue/picture for all items.
3. It shall be composed of 3 clear pictures minimum
4. The comparison table to the technical specification should be supplied

### **MANAGEMENT REQUIREMENTS**

1. **Warranty:** Minimum two (2) years manufacturer warranty
2. **Catalogue:** Colored catalogue with at least 3 clear images per item (front, side, control panel)
3. **Technical Compliance Table:** Must be provided mapping offered items against RFQ requirements
4. **Certificates & Approvals:** Equipment must be accompanied by all copies of safety, standards, and quality certificates and all manufacturer approval/certificates

### **INSTRUCTIONS TO BIDDERS**

#### **Eligibility Documents (mandatory, to be included in the technical proposal):**

- Valid RDB Registration Certificate (field: Medical Equipment Supply)
- Valid Tax Clearance Certificate (RRA)
- Valid RSSB Certificate
- RDB Non-Bankruptcy Certificate
- Bid Security of 10% of the total amount
- Certificate of Good Completion for at least five (5) similar tenders executed in the past five (5) years (attach award letters or completion certificates)
- Delivery schedule

#### **Quotation Format:**

#### **Two-envelope system:**

**Technical Proposal** (eligibility documents, Products Catalogue, technical specifications, compliance table, warranty terms, delivery schedule)

#### **Financial Proposal** (price, delivery and installation charges, Quotation validity)

- Prices must be quoted in Rwandan Francs (RWF), all-inclusive (CIF Kigali). Any price escalation clauses must be clearly stated.
- Validity: Quotation must remain valid for at least 90 days from the submission deadline.

- Delivery & Installation: All equipment to be delivered, installed, and commissioned at the hospitals.

**Submission Deadline:**

Submit your signed and stamped quotations in two separate sealed envelopes (Technical and Financial) no later than **Friday, 13th/06/2025, 5:00 PM Kigali time**. Late submissions will be rejected.

**Submission Address:**

To: Senior Finance & Operations Manager

Jhpiego Rwanda Country Office

Gasabo, Kacyiru, KG 5 St Road, House Number 48

P.O. Box 6952, Kigali, Rwanda

Tel: (+250) 788 381 188

Email for inquiries: [Rwanda.Procurement@jhpiego.org](mailto:Rwanda.Procurement@jhpiego.org)

All inquiries must be submitted in writing at least two (2) business days before the submission deadline.

**Note:** Jhpiego reserves the right to accept or reject any quotation, to annul the bidding process, and to reject all proposals at any time prior to contract award, without thereby incurring any liability to the affected bidder(s) or any obligation to inform the affected bidder(s) of the grounds for Jhpiego's action.

**Done on 05 June 2025**

**Valens Ndonkeye**

Sr. Finance and Operations Manager

Jhpiego Rwanda Country Office

Tel: (+250) 788 381 188