

## Contract | अनुबंध



Contract No | अनुबंध क्रमांक: GEMC-511687738076142

Generated Date | अनुबंध तिथि: 05-Jan-2024

Organisation Details   संगठन विवरण	Buyer Details   खरीदार विवरण
Type   प्ररूप : Central Autonomous Ministry   मंत्रालय : Ministry of Education Department   विभाग : Department of School Education and Literacy Organisation Name   संगठन का नाम : Navodaya Vidyalaya Samiti Office Zone   कार्यालय क्षेत्र : EAST	Designation   पद : SMT Contact No.   संपर्क नंबर : - Email ID   ईमेल आईडी : buyer150.nvs.bh@gembuyer.in GSTIN   जीएसटीआईएन : - Address   पता : JAWAHAR NAVODAYA VIDYALAYA, NAWANAGAR, BUXAR, BIHAR-802129, India

Financial Approval Detail   वित्तीय स्वीकृति विवरण	Paying Authority Details   भुगतान प्राधिकरण विवरण
IFD Concurrence   आईएफडी सहमति : No Designation of Administrative Approval   प्रशासनिक अनुमोदन का पदनाम : PRINCIPAL Designation of Financial Approval   वित्तीय अनुमोदन का पदनाम : PRINCIPAL	Role : PAO Payment Mode   भुगतान का तरीका : Offline Designation   पद : UDC Email ID   ईमेल आईडी : pao87.nvsn.bh@gembuyer.in GSTIN   जीएसटीआईएन : - Address   पता : JAWAHAR NAVODAYA VIDYALAYA, NAWANAGAR, Buxar, BIHAR-802129, India

Seller Details   विक्रेता विवरण	
GeM Seller ID   जेम विक्रेता आईडी : F1DE180000676163 Company Name   कंपनी का नाम : INDIAN ARMY & POLICE STORE Contact No.   संपर्क नंबर : 07991136251 Email ID   ईमेल आईडी : indianarmyandpolicestore@gmail.com Address   पता : HOTWASI, MAHUATOLI, BALALONG,,NAGRI, DHURWA, Ranchi, JHARKHAND-834004, - MSME Registration number   एमएसएमई पंजीकरण संख्या : - GSTIN   जीएसटीआईएन : 20GMNPS2103F1ZD	

\*GST / Tax invoice to be raised in the name of | जिसके नाम के पक्ष में GST/TAX इनवॉइस पेश किया जाएगा - Buyer

Delivery Instructions | वितरण निर्देश : NA

Product Details   उत्पाद विवरण						
#	Item Description   आइटम विवरण	Ordered Quantity   आइटम विवरण	Unit   इकाई	Unit Price (INR)   इकाई मूल्य (INR)	Tax Bifurcation (INR)   कर विभाजन (INR)	Price (Inclusive of all Duties and Taxes in INR)   मूल्य (INR में सभी शुल्क और कर सहित)
1	<b>Product Name</b>   उत्पाद का नाम : Unbranded RO Water Plant for Hemodialysis 500 Ltrs/Hr Capacity <b>Brand</b>   ब्रांड : NA <b>Brand Type</b>   ब्रांड प्रकार : Unbranded <b>Catalogue Status</b>   कैटलॉग की स्थिति : Catalogue not verified by OEM <b>Selling As</b>   कैसे बेचा जा रहा है : Reseller not verified by OEM <b>Category Name &amp; Quadrant</b>   श्रेणी का नाम और चतुर्थांश : RO Water Plant for Hemodialysis (Q3) <b>Model</b>   मॉडल : Eureka50002 <b>HSN Code</b>   एचएसएन कोड : HSN not specified by seller	1	pieces	465,000	NA	465,000
Total Order Value   कुल ऑर्डर मूल्य (in INR)						465,000

Consignee Detail   परेषिती विवरण						
S.No   क्र.सं.	Consignee   परेषिती	Item   वस्तु	Lot No.   लॉट नंबर	Quantity   मात्रा	Delivery Start After   दिनांक के बाद डिलीवरी शुरू करना है	Delivery To Be Completed By   वितरण पूरा कब तक करना है
1	Designation   पद : Vice Principal Email ID   ईमेल आईडी : buyer10.nvsn.bh@gembuyer.in Contact   संपर्क : - GSTIN   जीएसटीआईएन : - Address   पता : JAWAHAR NAVODAYA VIDYALAYA, NAWANAGAR, BUXAR, BIHAR-802129, India	Unbranded RO Water Plant for Hemodialysis 500 Ltrs/Hr Capacity	-	1	05-Jan-2024	20-Jan-2024

**Product Specification for Unbranded RO Water Plant for Hemodialysis 500 Ltrs/Hr Capacity**

Specification   विनिर्देश	Sub-Spec   उप-विनिर्देश	Value   मूल्य
GENERAL FEATURES	Clinical Purpose	RO water for hemodialysis to be applied in therapies such as HD, SLED, HDF, HF
	Product name	RO Water Plant for Hemodialysis
	Plant capacity (Ltrs/Hr)	500
	Plant type	Microprocessor/microcontroller controlled dual stage RO water system ( Feed water TDS more than 500 PPM)
	Various stages in RO water treatment process	Pre-filter or Sediment filter, Carbon filter, Water softner, Reverse Osmosis system, Distribution system, Storage Tank, Ultraviolet disinfection, Final filtration, Final alarms, System monitoring and disinfection
	System type	Offline
	System shall have microprocessor based control panel for ease of operation and inbuilt process logic	Yes
	System shall have auto start/stop based on water level in the supply tank	Yes
	System shall be equipped with Electrical panel/MCB for plant protection	Yes
PRE-TREATMENT	Raw water mesh filtering size to prevent big dust/sand particles with back wash control	50 to 200 microns
	Vertical water tank with automatic water float	Yes
	Raw water pump with automatic pressure control and run dry protection	Yes
	Raw water feed pump type	Centrifugal
	Double raw water pump ( one active and the other stand by)	Yes
	Raw water feed pump capacity (LPH)	2000
	Sand/iron filter with 20micron filter and sample valve	Yes
	Softener for one in operation while the other in standby and vice-versa with sample valve with brain tank and automatic regeneration capacity	Yes
	Double activated Carbon filter to remove Chlorine and Chloramines with sample valve	Yes
	Micron particle filter after activated Carbon filter for removing suspended particles more than micron size	5
	All pretreatment module should be programmable back wash and regeneration facility	Yes
	Online rotameter for measuring flow	Yes
	Pressure monitoring facility of all filtering stages with glycerin filled SS pressure gauge	Yes
Sample valve facility for all filtering stages	Yes	
RO UNIT FOR MAIN TREATMENT	Distribution of RO plant water sufficient for number of dialysis machines	30-40
	Monitoring facility for input and output water conductivity, feed water pressure and injection flow rate	Yes
	Alarm against low feed water, high output conductivity and high temperature of pump motor	Yes
	Unused water feed back facility to RO unit for saving on water injection in online mode	Yes
	The unit should be programmable and automatic rinsing/flushing facility at regular intervals when system is not in use, to prevent drying of filter media	Yes
	Emergency mode operation to run permeate output in case of electronic failure	Yes
RO MEMBRANE	Efficiency of unit with maximum saving of water (%)	55 to 75
	Sample valve for permeate	Yes
	Material of Booster pump (offline)	NA
	Vertical storage water tank with automatic float control for filling and dry run protection for off-line system	Yes
	Direct and indirect feed for full close loop distribution piping system	Yes
	Chemical disinfection and decalcification	Semi-automatic

POST TREATMENT	Provision of volume controlled heat disinfection of permeate loop complied with ISO 15883	Yes
	The permeate should be supplied to distribution loop through 0 point 2 micron bacteria filter for offline system using pex piping and high grade SS316 push pull connectors	Yes
	The distribution loop should contain loop pressure regulator to maintain the desired loop back pressure	Yes
USER INTERFACE	Digital display of values of Conductivity/permeable flow/reject flow/temperature/pressure monitoring	Yes
	Touch screen/button type illuminated display for easy to finger operation of user	Yes
	Signal and control lights with possibility to connect external lights	Yes
	Software update via SD Card or Ethernet	No
	Memory storage device	No
STANDARD CONSUMABLES AND ACCESSORIES	All media and consumables ( membrane, cartidge filter, filter media, chemicals, softner salt) for setting up and standardization should be provided	Yes
	Seller should ensure the availability of all consumables required to run the RO plant during complete warranty period to be purchased by the buyer independently at actual cost	Yes
	UPS of Suitable Capacity to pe provided	No
CERTIFICATIONS & REPORTS	Submission of Test Report for permeate water quality as per AAMI/ISO-13959/13500 from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification after installation	YES
	Product certification	BIS
	Certificate No	15
	Certificate Date	20
	Certificate issuing authority	25
	Four digit number of notified body If product is EU-CE certified	4
	Certification, performance and safety standards specific to the device	IS:13450(Part 1)
	Submission of all the certifications and test reports to the buyer along with supplies on demand	Yes
PRE-INSTALLATION REQUIREMENT FOR THE BUYER	All Pre-installation requirements along with site preparation should be provided by the seller to the buyer well in advance before the supply of equipment	Yes
	All Pre-installation requirements should be done by the buyer	Yes
	Minimum room size required for the plant (H x W x L) m	11X20X20
	Feed water line till RO room	Yes
	Number of drain outlets required	3
	Overhead tank for storage depending on plant size	Yes
	Door width required for RO room (cm)	305
	Airconditioning as per rquirements	Yes
	Power requirements	220-240 VAC, single phase
INSTALLATION & TRAINING	Supplier to perform installation, safety and operation checks before handover	Yes
	Training of users in operation and basic maintenance shall be provided	Yes
WARRANTY & MAINTENANCE	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	2
	User technical and maintenance manual detailing complete maintaining schedule with routine maintenance should be provided	Yes
	Contact details of manufacturer, supplier and local service agent to be provided	Yes
<p><b>Note   टिप्पणी:</b> Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.</p>		

NA

## Terms and Conditions | नियम और शर्तें

### 1. Special terms and conditions- Version:2 effective from 24-11-2023

#### 1.1 1. Comprehensive warranty:

Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

#### 2. Service centres:

Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

#### 3. Source of supply:

It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.

#### 4. Packing and Marking:

Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.

#### 5. Spare Parts:

Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied

#### 6. Installation, Training, Manuals:

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

#### 7. Electrical safety checking:

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

#### 8. Software:

All software updates should be provided free of cost during warranty period.

### 2. General Terms and Conditions-

2.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.

2.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।