

## अनुबंध | Contract



अनुबंध क्रमांक | Contract No: GEMC-511687730857868

अनुबंध तिथि | Generated Date : 01-Feb-2024

बोली/आरए/पीबीपी संख्या | Bid/RA/PBP No.: [GEM/2023/B/3898169](#)

| संगठन विवरण   Organisation Details  | खरीदार विवरण   Buyer Details  |
|---|---|
| प्रकार   Type : State Government<br>मंत्रालय   Ministry : -<br>विभाग   Department : Information Technology Department Lakshadweep<br>संगठन का नाम   Organisation Name : N/A<br>कार्यालय क्षेत्र   Office Zone : Lakshadweep | पद   Designation : Buyer Hospital<br>संपर्क नंबर   Contact No. : 04896-262209-<br>ईमेल आईडी   Email ID : bfonlep-lk@gov.in<br>जीएसटीआईएन   GSTIN : -<br>पता   Address : Department of Health, Indira Gandhi Hospital, Near Post office Junction, LAKSHADWEEP, LAKSHADWEEP-682555, India |

| वित्तीय स्वीकृति विवरण   Financial Approval Detail   | भुगतान प्राधिकरण विवरण   Paying Authority Details  |
|--|--|
| आईएफडी सहमति   IFD Concurrence : No<br>प्रशासनिक अनुमोदन का पदनाम   Designation of Administrative Approval : Advisor to the Honorable Administrator<br>वित्तीय अनुमोदन का पदनाम   Designation of Financial Approval : Advisor to the Honorable Administrator | Role: PAO<br>भुगतान का तरीका   Payment Mode: Offline<br>पद   Designation : AO<br>ईमेल आईडी   Email ID : lak-ao@nic.in<br>जीएसटीआईएन   GSTIN : N<br>पता   Address : LITSS, Dept. of IT 1st Floor New Secretariat Building, LAKSHADWEEP, LAKSHADWEEP-682555, India |

| विक्रेता विवरण   Seller Details  |
|--|
| जेम विक्रेता आईडी   GeM Seller ID : 45C4180000360303<br>कंपनी का नाम   Company Name : TRANSASIA BIO-MEDICALS LIMITED<br>संपर्क नंबर   Contact No. : 09321413273<br>ईमेल आईडी   Email ID : tenders@transasia.co.in<br>पता   Address : 8, CHANDIVALI STUDIO ROAD, ANDHERI EAST, MUMBAI, Maharashtra-400072, -<br>एमएसएमई पंजीकरण संख्या   MSME Registration number : -<br>जीएसटीआईएन   GSTIN : 27AAACT2038C1ZT |

\*जिसके नाम के पक्ष में GST/TAX इनवॉइस पेश किया जाएगा | GST / Tax invoice to be raised in the name of - Buyer

वितरण निर्देश | Delivery Instructions : NA

| #  | आइटम विवरण   Item Description   | आइटम विवरण   Ordered Quantity | इकाई   Unit | इकाई मूल्य (INR)   Unit Price (INR) | कर विभाजन (INR)   Tax Bifurcation (INR) | मूल्य (INR में सभी शुल्क और कर सहित)   Price (Inclusive of all Duties and Taxes in INR) |
|--|---|-------------------------------|-------------|-------------------------------------|---|---|
| 1  | उत्पाद का नाम   Product Name : ERBA both (automatic and manual) Hematology Analyser ( 3 PART )<br>ब्रांड   Brand : ERBA<br>ब्रांड प्रकार   Brand Type : Registered Brand<br>कैटलॉग की स्थिति   Catalogue Status: OEM verified catalogue<br>कैसे बेचा जा रहा है   Selling As : OEM<br>श्रेणी का नाम और चतुर्थांश   Category Name & Quadrant : Automated Hematology Analyser (Part 3) (Q3)<br>मॉडल   Model: H 360 with 3 Year Warranty<br>एचएसएन कोड   HSN Code: 90278990 | 4                             | pieces      | 145,000                             | NA                                      | 580,000   |
| 2  | उत्पाद का नाम   Product Name : ICT charges for Automated Hematology Analyser (Part 3)<br>ब्रांड   Brand : NA<br>ब्रांड प्रकार   Brand Type : Unbranded<br>कैटलॉग की स्थिति   Catalogue Status: Catalogue not verified by OEM<br>कैसे बेचा जा रहा है   Selling As : Reseller not verified by OEM<br>श्रेणी का नाम और चतुर्थांश   Category Name & Quadrant : Addon Services for bid (Q3)<br>मॉडल   Model: ICT<br>एचएसएन कोड   HSN Code: HSN not specified by seller       | 4                             | pieces      | 43,800                              | NA                                      | 175,200   |
| कुल ऑर्डर मूल्य   Total Order Value (in INR) |   |                               |             |                                     |   | 755,200   |

| परोक्षिती विवरण   Consignee Detail |
|------------------------------------|
| दिनांक के बाद वितरण पूरा कब        |

| क्र.सं.   S.No | परोक्षता   Consignee  | वस्तु   Item   | लॉट नंबर   Lot No. | मात्रा   Quantity | डिलीवरी शुरू करना है   Delivery Start After | तक करना है   Delivery To Be Completed By |
|----------------|---|--|--------------------|-------------------|---|--|
| 1              | पद   Designation :-<br>ईमेल आईडी   Email ID : smshihab-lk@gov.in<br>संपर्क   Contact :-<br>जीएसटीआईएन   GSTIN :-<br>पता   Address : Lakshadweep Administration Office Indira Gandhi Road,<br>Willindon Island Kochi,<br>ERNAKULAM, KERALA-682003, India | ERBA both (automatic and manual)<br>Hematology Analyser ( 3 PART ) | -                  | 4                 | 01-Feb-2024                                 | 16-Feb-2024                              |

**Product Specification for ERBA both (automatic and manual) Hematology Analyser ( 3 PART )**

| विनिर्देश   Specification           | उप-विनिर्देश   Sub-Spec                                 | मूल्य   Value  |
|-------------------------------------|---|--|
| Standards                           | Certifications for the product                          | CE European (with verified number)   |
|                                     | Certification number and date                           | PICL/CE/0817/3970 Dated: 10.08.17  |
|                                     | Conformity to quality management standards              | ISO 9001 & ISO 13485   |
|                                     | Conformity to electrical safety standards               | IEC 61010-1,IEC 61010-2-081,IEC 61010-2-101 or BIS equivalent  |
| Performance Parameters              | UMDNS Name and Code                                     | "Analyzers, Laboratory, Hematology, Cell Counting, Automated [17-741],Hematological cell analyser Analyzers, Laboratory, Hematology, Cell Counting, Semiautomated [17-742] Analyzers, Point-of-Care, Whole Blood, Hematology [18-513] Computer Aided Detection Systems, Pathology Slide, Blood Cell Classification [24-248]" |
|                                     | Clinical Purpose  | Hematology Analyzers are used to count and characterize blood cells for disease detection and monitoring. Basic analyzers return a complete blood count(CBC) with a three-part differential white blood cell(WBC) count.   |
|                                     | Type of Configuration                                   | Bench top  |
|                                     | Type of system offered                                  | closed system  |
|                                     | Type of Automation                                      | Fully Automatic  |
|                                     | Automatic Start Up,shut down and sample analysis        | Yes  |
|                                     | Analysis principle                                      | Based on principle of counting and sizing  |
|                                     | Multi channel analysis for better resolution            | Yes  |
|                                     | Type of cell counting                                   | 3 part WBC differential  |
|                                     | Testing mode selection available                        | CBC Mode   |
|                                     | Analysis available                                      | WBC,Lymph#,Lymph%, Granulocytes #,Granulocytes%Mixed #, Mixed %,RBC,HGB,HCT,MCV,PCV,MCH,MCHC,RDW-SD,RDW-CV,PLT,MPV,PDW-SD,PDW-CV,PCT,PLCR ,P-LCC   |
|                                     | Should have three histograms of WBC,RBC,PLT             | Yes  |
|                                     | MID% analysis   | Yes  |
|                                     | MTD# analysis   | Yes  |
|                                     | Gran% analysis  | Yes  |
|                                     | Gran# analysis  | Yes  |
|                                     | Analysis method for WBC                                 | electrical Impedance   |
|                                     | Method for platelet measurement                         | electrical Impedance   |
|                                     | RBC Measurement method                                  | electrical Impedance   |
|                                     | Hb measurement  | Cyanide free Colorimetry   |
|                                     | Types of modes of running sample                        | Open vial,Pre dilute   |
|                                     | Maximum sample aspiration volume needed in any of modes | should be less than 50 µl  |
|                                     | Minimum sample volume required                          | should be 20 µl or less  |
|                                     | Throughput capacity of analyser in (samples/ hour)      | 50-60  |
|                                     | Linearity of Platelet                                   | " 0 to 300 * 10 <sup>3</sup> cells/ micro litres"  |
|                                     | RBC Linearity   | 0 to 8 x 10 <sup>6</sup> per micro litre or more   |
| Hemoglobin linearity                | 0 to 25 gm per litre                                    |  |
| WBC linearity                       | " 0 to 3000 * 10 <sup>3</sup> cells/ micro litres"      |  |
| Analyses time for cytopenic samples | Yes   |  |
| Directly measures MCV               | Yes   |  |

|  |   |  |
|--|---|--|
|  | Time taken by the analyser to produce the test results(Analysis time) in seconds  | 40-60  |
|  | Availability of Auto dilution   | Yes  |
|  | Types of reagents   | 1 hemolyzing reagent,1 diluent,1 cleansing solution  |
|  | Quality assurance system with calibration and controls  | Yes  |
|  | Atleast 3 quality control programs  | Yes  |
|  | Type of Calibration   | both (automatic and manual)  |
|  | Direct aspiration for capillary blood from finger prick   | Yes  |
|  | Floating discriminator for platelets and RBC counting for reliable RBC and PLT data   | Yes  |
|  | Automatic probe wipe  | Yes  |
|  | Separate diluting nozzles for RBC and WBC   | Yes  |
|  | Double bathing mechanism  | Yes  |
|  | Automatic electric clog removal   | Yes  |
| Data Management and Display              | Type of data management   | Inbuilt system   |
|  | Display   | LED  |
|  | Inbuilt monitor size in inches  | More than 5  |
|  | PC Monitor size in inches (When PC provided externally)   | NA (if Inbuilt system)   |
|  | PC hard disk  | NA   |
|  | RAM capacity of PC System   | NA (if no PC provided)   |
|  | Processor   | NA   |
|  | HIS/LIS Interface   | RS232  |
|  | Type of external storage  | USB  |
|  | No of USB Port  | 2  |
|  | Data management systems   | Provide histograms in display and print  |
|  | Facility for user defined flagging  | Yes  |
|  | Type of user Interface or data entry  | all three (touchscreen, handheld barcode reader facility and manual)   |
|  | "Database capability of storing sets of results and graphics"   | " >=5000 to 10000"   |
|  | Facility for workload recording   | Yes  |
|  | "Auto stop function in event of unacceptable control data"  | Yes  |
|  | Ability to transmit results to host computer  | Yes  |
|  | Have auto cleaning function in the analyser's software  | Yes  |
|  | Printer facility  | Yes  |
| Type of printer unit                     | Inbuilt   |  |
| Printer type                             | Thermal printer   |  |
| Power Requirements                       | Type of power supply in VAC,HZ  | 230-240 VAC,50-60 Hz   |
|  | Power Backup facility   | Yes  |
|  | Type of UPS   | Online   |
|  | Rating of UPS in KVA  | 1 KVA  |
|  | Back up time in minutes   | 30 min   |
| Accessories, spare parts and consumables | "Offered equipment unit to be supplied with sufficient consumables (with at least 2/3rd of total shelf life) required for, sufficient to carry out haematological testing of samples" | 500  |
|  | Two set of all tubings  | Yes  |
|  | Reagent expiry time should be minimum of 1 year   | Yes  |
|  | Alerts for operator for level of reagents and to empty waste when indicated   | Yes  |
|  | Availability of micro capillary adapter   | Yes  |
|  | Operating temperature and humidity  | Capable of operating continuously in ambient temperature of 15 to 35 deg C and relative humidity of 15 to 85% in ideal circumstances |
|  | Warranty (Option of comprehensive warranty is available   |  |

|  |  |             |
|--|--|-------------|
| Miscellaneous Parameters                         | through bidding only, which if opted will supersede normal warranty in the catalogue)  | 3           |
|  | Number of installations in Central /State/PSU Govt Hospitals ( Hint: Seller should supply a performace certificate of the device to the buyer incase demanded after placement of order)  | More than 3 |
|  | OEM/Reseller ( if supplied by reseller shall ensure uninterputed availability of all spares for 10 years)  | Yes         |
|  | Availability of toll free facility for technical support maintained by OEM or authorized agencies  | Yes         |
|  | During the warranty period it shall be ensured that all the break down calls are attended with in 48 hrs and the complete details of service agents / contact details should be furnished to buyer and consignee at time of supplies | Yes         |
|  | User/Technical/Maintenance manuals to be supplied in English in hard and soft copy   | Yes         |
|  | Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished  | Yes         |
|  | List of important spares and accessories, with their part numbers to be supplied to the buyer at the time of supplying the equipment   | Yes         |
|  | Installation and demonstration of equipment and training to be provided after completing supplies before acceptance  | Yes         |
|  | The Principal Manufacturer must have direct Presence/approved service center In India  | Yes         |
|  | Copies of reports and certifications to be furnished to buyer on demand at time of supplies  | Yes         |
| Calibration certificates as per NABH requirement | Yes  |             |

**उपरोक्त मद के लिए स्थापना कमीशनिंग और परीक्षण (आईसीटी) विवरण | Installation Commissioning and Testing (ICT) details for the above item::**

|   |         |
|---|---------|
| % of Product Cost Payable on Product Delivery                               | 70 %    |
| Min Cost Allocation for ICT as a % of product cost                          | 30 %    |
| Number of days allowed for ICT after site readiness communication to seller | 30 Days |

**शुद्धिपत्र | Corrigendum**

1. तक बढ़ाया गया | Extended Upto : 2023-09-19 14:00:00

**ईपीबीजी विवरण | ePBG Detail**

|   |    |
|---|----|
| सलाहकार बैंक   Advisory Bank :            | NA |
| ईपीबीजी प्रतिशत (%)   ePBG Percentage(%): | NA |

**नियम और शर्तें | Terms and Conditions**

**1. Special terms and conditions- Version:2 effective from 21-10-2022**

- 1.1
- All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation (CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
  - The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
  - The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
  - Comprehensive warranty: Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  - Service centres: Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of

- service requests also to be provided buyer/consignee at the time of supplies.
6. Source of supply: It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  7. Packing and Marking: Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  8. Spare Parts: Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
  9. Installation, Training, Manuals: Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
  10. Electrical safety checking: Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
  11. Software: All software updates should be provided free of cost during warranty period.
  12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

## 2. General Terms and Conditions-

- 2.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.
- 2.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

## 3. Buyer Added Bid Specific Terms and Conditions-

### 3.1 Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

### 3.2 Scope of Supply:

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.