

अनुबंध | Contract



अनुबंध क्रमांक | Contract No: GEMC-511687710621906

अनुबंध तिथि | Generated Date : 01-Feb-2024

बोली/आरए/पीबीपी संख्या | Bid/RA/PBP No.: [GEM/2023/B/3627205](#)

संगठन विवरण Organisation Details	खरीदार विवरण Buyer Details
प्रकार Type : Central Government मंत्रालय Ministry : Ministry of Health and Family Welfare विभाग Department : Department of Health Research संगठन का नाम Organisation Name : N/A कार्यालय क्षेत्र Office Zone: Bhopal Memorial Hospital & Research Centre Bhopal	पद Designation : Purchase Officer संपर्क नंबर Contact No. : 0755-2734561- ईमेल आईडी Email ID : buyer2.dhr.bpl@gembuyer.in जीएसटीआईएन GSTIN : 23AAEAT4818Q3ZY पता Address : Raisen By Pass Road Karond Bhopal, BHOPAL, MADHYA PRADESH-462038, India

वित्तीय स्वीकृति विवरण Financial Approval Detail	भुगतान प्राधिकरण विवरण Paying Authority Details
आईएफडी सहमति IFD Concurrence : No प्रशासनिक अनुमोदन का पदनाम Designation of Administrative Approval: Director Incharge वित्तीय अनुमोदन का पदनाम Designation of Financial Approval: Director Incharge	Role: PAO भुगतान का तरीका Payment Mode: Offline पद Designation : Computer Operator2 ईमेल आईडी Email ID : pao1.dhr.mp@gembuyer.in जीएसटीआईएन GSTIN : - पता Address: Raisen By Pass Road Karond Bhopal, BHOPAL, MADHYA PRADESH-462038, India

विक्रेता विवरण Seller Details	
जेम विक्रेता आईडी GeM Seller ID : AA3318000090783 कंपनी का नाम Company Name : ARYAN DISTRIBUTORS संपर्क नंबर Contact No. : 09425007273 ईमेल आईडी Email ID : aryandistributor@yahoo.in पता Address : B - 110, HABIBGANJ, ALKAPURI, Bhopal, MADHYA PRADESH-462024, - एमएसएमई सत्यापित MSME verified : No एमएसएमई पंजीकरण संख्या MSME Registration number : UDYAM-MP-10-0022022 एमएसएमई सामाजिक श्रेणी MSE Social Category : General एमएसएमई लिंग श्रेणी MSE Gender : Female जीएसटीआईएन GSTIN: 23BMIPS8854D1ZR	

*जिसके नाम के पक्ष में GST/TAX इनवॉइस पेश किया जाएगा | GST / Tax invoice to be raised in the name of - Buyer

वितरण निर्देश | Delivery Instructions : The above procurement is under buyback and Buyback value offered by the bidder in cost breakup format Rs. 35000/- will be deducted from above ordered value.

उत्पाद विवरण Product Details						
#	आइटम विवरण Item Description	आइटम विवरण Ordered Quantity	इकाई Unit	इकाई मूल्य (INR) Unit Price (INR)	कर विभाजन (INR) Tax Bifurcation (INR)	मूल्य (INR में सभी शुल्क और कर सहित) Price (Inclusive of all Duties and Taxes in INR)
1	उत्पाद का नाम Product Name : Olympus Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic) ब्रांड Brand : Olympus ब्रांड प्रकार Brand Type : Registered Brand कैटलॉग की स्थिति Catalogue Status: OEM verified catalogue कैसे बेचा जा रहा है Selling As : OEM verified Reseller श्रेणी का नाम और चतुर्थांश Category Name & Quadrant : Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic) (Q3) मॉडल Model: EVIS EXERA III - CV190 एचएसएन कोड HSN Code: HSN not specified by seller	1	pieces	6,935,890	NA	6,935,890
कुल ऑर्डर मूल्य Total Order Value (in INR)						6,935,890

परोक्ष विवरण Consignee Detail						
क्र.सं. S.No	परोक्ष Consignee	वस्तु Item	लॉट नंबर Lot No.	मात्रा Quantity	दिनांक के बाद डिलीवरी शुरू करना है Delivery Start After	वितरण पूरा कब तक करना है Delivery To Be Completed By

1	पद Designation : - ईमेल आईडी Email ID : con11.dhr.bhopal@gembuyer.in संपर्क Contact : 755-2742212-5510 जीएसटीआईएन GSTIN : - पता Address : Raisen By Pass Road Karond Bhopal, BHOPAL, MADHYA PRADESH-462038, India	Olympus Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic)	-	1	01-Feb-2024	01-May-2024
---	--	---	---	---	-------------	-------------

Product Specification for Olympus Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic)

विनिर्देश Specification	उप-विनिर्देश Sub-Spec	मूल्य Value
GENERAL FEATURES	Product description	Video endoscope unit having Upper Gastrointestinal Scope
	Clinical purpose	To diagnose and sometimes treat conditions that affect the upper part of patient's digestive system, including the esophagus, stomach and beginning of the small intestine (duodenum)
	System shall include	Gastrovideoscope, Video Processor, Computer System with High Definition Medical Grade Monitor(compatible with the scope), Printer, Trolley , endoscopic software and Standard Accessories
GASTROVIDEOSCOPE SPECIFICATION	Purpose of Gastrovideoscope	Both Diagnostic and Therapeutic
	Patient Application	Adult
	Gastrovideoscope should be fully immersible in disinfectant solution with or without resistant cap	Yes
	Type of inbuilt image sensor	CCD/CMOS
	Image sensor should be HDTV Compatible	Yes
	Provision of near/close focus observation	Yes
	Provision of image enhanced endoscope technology or equivalent chrome endoscopy imaging for detailed mucosal study	Yes
	Provision of inbuilt identification memory chip for monitor display of scope model no , serial no, white balancing memory No of connections	Yes
	Provision of forward/auxillary water jet for mucosal cleaning	Yes
	Field of view/ Angle of view for Normal/Near Focus	140 degree
	Direction of view	Forward viewing
	Depth of Field	2 mm to 100 mm or better
	Distal end outer Diameter	9.2 millimeter
	Insertion tube outer diameter	9.2 millimeter
	Instrument channel inner diameter	2.8 mm or more
	Working Length of Gastrovideoscope	1030 millimeter
	Total Length of Gastrovideoscope	1350 millimeter
	Minimum visible distance	3.0 mm or closer from the distal end
	Minimum Tip Bending range (Up)	210 degree
	Minimum Tip Bending range (down)	90 degree
	Minimum Tip Bending range (Right)	100 degree
	Minimum Tip Bending range (Left)	100 degree
	The surface of the gastroscope shall be free of pores, cracks, and remainders of tooling agents	Yes
Materials used for the outer surface of the insertion portion shall be evaluated for biological compatibility	Yes	
	Video Processor should be compatible with the gastrovideoscope	Yes
	Video processor should be compatible with digital HD-SDI / DVI output for HDTV monitor and should contain electronics for clear visibility of near and far objects	Yes
	Provision of RGB, S-video, composite video output function	Yes
	Equipped With high resolution HDTV imaging capacity (1080 P)	Yes
	Provision of chromo endoscopy imaging	Yes
	Provision of Portable memory and USB slot for still image recording	Yes
	Provision of automatic IRIS control and white balance	Yes
	Provision of Picture in picture display and index function ability	Yes
	Provision of Electronic Zoom upto	1.5 X

VIDEO PROCESSOR SPECIFICATIONS	Equipped with memory backup for settings and lithium battery	Yes
	Provision of prefreeze function for image stabilization	Yes
	Provision of Light Source	Separate
	Light Source Type	Xenon lamp
	Power of Lamp (watt)	300
	Provision of Standby/Emergency Lamp of equal rating	Yes for Xenon Lamp
	Minimum Average lamp life	500 hours of continuous use for Xenon lamp
	Provision for lamp switch over from main lamp to stand by lamp in case of main lamp failure	Yes
	Provision of backlit front panel indicator	Yes
	Equipped with automatic light adjustment facility	Yes
	Cooling Method	Forced air cooling
	Equipped with regulated air feeding pump and fan with low noise	Yes
	Video processor should be supplied with minimum 2 KW online UPS	Yes
	Provided with two water bottles compatible with processor	Yes
MONITOR SPECIFICATIONS	Monitor should be compatible with the scope	Yes
	Monitor should be high definition medical grade	Yes
	Type of Monitor	Color LCD
	Minimum Size of Monitor	24 inch
	Monitor should have aspect ratio of 16:9/16:10 with output of (1080/601:NTSC) (1080/501:PAL)	Yes
	Monitor should have provision of accepting HD-SDI and DVID signal	Yes
	Monitor should have low power consumption	Yes
	Provision of picture in picture and picture out picture for viewing side by side split screen images	Yes
	Monitor should have wide viewing angle	Yes
COMPUTER SYSTEM SPECIFICATIONS	Processor	core i3
	Operating System	Window 8 or above
	RAM Size	4 gigabyte
	Hard Disk Capacity	1 terabyte
ENDOSCOPIC SOFTWARE SPECIFICATIONS	High Definition patient reporting software to record and capture the still images during the procedure should be provided with the unit	Yes
	Provision of Foot switch and image capture dongle along with the software	Yes
	Image capture software should be compatible with the computer system and Printer provided	Yes
PRINTER SPECIFICATIONS	Printing Technology	Laser
	Type of Printer	Color
	USB connectivity in Printer	Yes
TROLLEY SPECIFICATIONS	Material of Trolley	Stainless Steel 304 Grade
	Trolley should have space to accommodate the video processor, monitor and all the required equipments	Yes
	Provision of Scope hanger in the trolley to hold the scope during the procedure	Yes
	Trolley should be powder coated good quality rust free and should be able to take load of minimum up to 70 Kg	Yes
	Trolley should have antistatic castor wheels having front locking facility to move the trolley	Yes
	Reusable Biopsy Forcep oval cup fenestrated	2
	Reusable Biopsy Forcep oval cup non fenestrated	2
	Hot Biopsy Forceps -Alligator cups with needle	Not Provided
	Hot Biopsy Forceps -Alligator cups without needle	Not Provided

STANDARD ACCESSORIES	Retrievable Basket	2
	Electrosurgical Snare	2
	Bipolar Probes	Not Provided
	Channel Cleaning Brushes	5
	Channel Opening Brushes	5
	Washing Pipe/Spray Catheter	Not Provided
	Injection needle (21 G)	10
	Grasping Forcep - rat tooth	4
	Reusable Rotable clip fixing device (Short and long)	Not Provided
	Reusable Extra suction and air water buttons	5 each
	Biopsy Channel Valve	Not Provided
	Extra bulb (as per light source)	Not Provided
	Leakage Tester	1
	All the accessories provided should be compatible with the endoscope and authorization certificate for compatibility by the parent company/supplier of the endoscope should be provided	Yes
	ELECTRICAL REQUIREMENTS	Input Power Supply
PACKING MODE	The complete unit should be packed including all the accessories in such a way that there is no transit damage takes place during transportation till destination	Yes
CERTIFICATIONS & REPORTS	Submission of Test Report/Quality Control Report to the buyer from parent manufacturer complying that the product meets the defined set of declared specifications	Yes
	Product Certification for the complete unit (Proof of the same to be submitted to buyer)	US-FDA,EU-CE
	Four digit number of notified body if product is EU-CE certified	0197
	ISO 13485 Certified Manufacturer (Proof of the same to be submitted to Buyer on demand)	Yes
	Gastroscope shall conform to IEC 60601-2-18 for safety and essential performance requirements (Proof of the same to be submitted to buyer on demand)	Yes
	Certification, performance and safety standards specific to the whole unit (Proof of the same to be submitted to buyer on demand)	IEC 60601-1
INSTALLATION & TRAINING	Supplier to perform installation, safety and operation checks before handover	Yes
	Onsite operational as well as general troubleshooting User level maintenance training shall be given to the user during supply and as when required by the user	Yes
WARRANTY & MAINTENANCE	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5
	User technical and maintenance manual detailing complete maintaining schedule with routine maintenance should be provided	Yes
	After sale support service facility to the user	Yes
	Contact details of manufacturer, supplier and local service agent to be provided	Yes

व्यापक रखरखाव के लिए शुल्क | Comprehensive maintenance charges for Olympus Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic)

सेवा वर्ष Service Year	सीएमसी प्रतिशत CMC Percentage	कर द्विभाजन Tax Bifurcation
CMC charges for 1 st year after warranty period in % of cost of equipment	6.5%	NA
CMC charges for 2 nd year after warranty period in % of cost of equipment	7%	
CMC charges for 3 rd year after warranty period in % of cost of equipment	8.3%	
CMC charges for 4 th year after warranty period in % of cost of equipment	9%	
CMC charges for 5 th year after warranty period in % of cost of equipment	9.5%	

खरीदार परिभाषित अतिरिक्त विशेषता के लिए | Buyer Defined Additional Specification for Olympus Video Endoscope Unit Having Upper Gastrointestinal Scope (Diagnostic and Therapeutic)

विनिर्देश Specification	मूल्य Value
Type of Monitor	Color LED
Technical Specifications of Lower GI Scope as per Annexure A mentioned in Buyer added bid specific ATC	Technical Specifications of Lower GI Scope as per Annexure A mentioned in Buyer added bid specific ATC
Technical Specifications of Side view Dudenovideoscope as per Annexure B mentioned in Buyer added bid specific ATC	Technical Specifications of Side view Dudenovideoscope as per Annexure B mentioned in Buyer added bid specific ATC

शुद्धिपत्र | Corrigendum

1. तक बढ़ाया गया | Extended Upto : 2023-07-29 16:00:00
2. GeM-Bidding-Corr-4961436-3.pdf: [यहाँ क्लिक करें | click here](#)

सीएमसी के लिए अतिरिक्त खंड | Additional Clauses for CMC

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

- CMC charges for 1st year after warranty period- Percentage to be indicated- A1
- CMC charges for 2nd year after warranty period- Percentage to be indicated- A2
- CMC charges for 3rd year after warranty period - Percentage to be indicated- A3
- CMC charges for 4th year after warranty period - Percentage to be indicated- A4
- CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

- 8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.
- 8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables
- (i) Number of years for which CMC required.
 - (ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$$C+C*((A1/100)/(1.10^n)+(A2/100)/(1.10^{n+1})+(A3/100)/(1.10^{n+2})+(A4/100)/(1.10^{n+3})+(A5/100)/(1.10^{n+4}))$$

and so on
C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5 shall depend on how many years CMC selected. For 3 year CMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

ईपीबीजी विवरण | ePBG Detail

सलाहकार बैंक Advisory Bank :	State Bank of India
ईपीबीजी प्रतिशत (%) ePBG Percentage(%):	3.00

बोली लगाने वाले को बोली के नियमों और शर्तों के अनुसार लागू ईपीबीजी प्रस्तुत करना होगा | The bidder shall furnish ePBG as applicable as per bid's terms and conditions

नियम और शर्तें | Terms and Conditions

1. General Terms and Conditions-

1.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.

1.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

2. Buyer Added Bid Specific Terms and Conditions-

2.1 Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.

2.2 Generic

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 3 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

2.3 Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

2.4 Generic

Experience Criteria: The Bidder or its OEM (themselves or through reseller(s)) should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

2.5 Generic

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

2.6 Scope of Supply:

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

2.7 OEM

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

2.8 Service & Support:

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

2.9 Service & Support:

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

2.10 Service & Support:

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

2.11 Certificates:

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

2.12 Certificates:

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

2.13 Warranty:

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

2.14 Warranty:

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope

Repair, Maintenance & Replacement of Spares without any additional cost

(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

2.15 Warranty:

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

2.16 Warranty:

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

2.17 Buyback for Non PFMS / Non GPA :

Buyback for Non PFMS/Non GPA:

1. GST will apply on the value of buyback and separate invoice is to be prepared by Government buyer/Consignee for buyback items offline outside the GeM system. Vendor should work out input credit for that GST applicable on the buy back while quoting the product/services.
2. Input GST Credit consideration by seller- seller shall offer the prices in the bid by the total amount of new goods including GST- total amount of old goods quoted by the firm excluding GST. This is because GST charged by buyer on buy-back value will be available for full ITC (input tax credit) on the hands of bidder. Therefore, for the purpose of quoting in the bid and evaluation, value as quoted by the bidder (excluding any GST on buy back) as indicated in column G below will be considered.

Illustrative examples for bidders are as under for understanding of the seller.

Excel sheet for explanation to Buyback ATC							
S.no	Cost of Fresh Good (A)	%age of GST on Fresh Good as considered by seller at the time of quote (B).	GST Amount on Fresh Good (C)	Cost of Buyback item (D)	%age of GST on Buyback items (E) as fixed by buyer in the bid	GST Amount on Buyback item (F) as applicable at the time of invoice by buyer	Quote of the Bidder (G={A+C-D})
1	100	28	28	20	8	1.6	108
2	100	9	9	20	8	1.6	89
3	110	10	11	24	8	1.92	97

Seller after the award of the GeM contract, will indicate all column A to G which will not alter their quoted price indicated in Colum-G

3. Successful lowest bidder has to provide detailed cost break up of tendered items as well as buy back items separately after award of contract through email to the government buyer/consignee.
4. The seller of new item /service provider shall deposit the cost of invoiced buy back item with GST to the buyer for allowing seller to lift up the material from the buyer/consignee premise.
5. Transportation of buy back items from buyer premises shall be arranged by the seller itself without any extra cost.
6. Government Buyer organisation /consignee will issue invoice of buy back items to the seller at the cost of buyback to be indicated to the buyer by the seller offline after receipt of order. This in no case will modify the bid in any manner.
7. The seller will be entirely responsible for the safe disposal of old items without affecting the environment in any manner as per prevailing statutory rules and Act/ law of the land.
8. The seller may visit the site and inspect the condition of buy back item in advance before quoting for the bid.
9. Receipt of payment by the buyer for Old items and invoicing by buyer will be out of GeM portal.
10. Government Buyer will upload the details of old goods along with the bid in the corrigendum with minimum price.
11. If the minimum price (Optional) is indicated and sellers of new item who are not agreeing to take back the old item at equal to or above that minimum price should not participate in the bid.
12. In case they participate into the bid and afterwards indicates the price of old goods, less than the minimum price shown by Government Buyer in corrigendum, EMD of the bidder will be forfeited.
13. Item Details Consignee wise will be indicated in the corrigendum like -
 1. Consignee wise
 2. Item Description
 3. Serial Number
 4. Model
 5. Brand
 6. Year of manufacturing
 7. Purchased on
 8. Age
 9. Size, Rating, Designation
 10. Quantity
 11. Min price (Optional)
 12. Unit
 13. Working Guidelines (/Functional/Non-Functional)
 14. Remarks
14. The successful bidder shall remove the old goods from the premise of the Government buyer/consignee at the time of the supply, installation and commissioning of fresh goods. Seller will be entitled for payment only after removal of old goods from Government buyer/consignee/user's premise and after supplying the fresh goods in terms of contract.
15. Note: This clause of ATC shall be operated by non PFMS; non GPA buyer i.e. all PSU working through their ERP and not integrated with GPA.

2.18 Forms of EMD and PBG:

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

Director, BMHRC
payable at
Bhopal

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

2.19 Buyer Added Bid Specific ATC:

Buyer Added text based ATC clauses

COMPREHENSIVE WARRANTY: - 5 YEARS to be submitted by OEM on Letterhead

Under Warranty:- Repair, Maintenance & Replacement of Spares without any additional cost

CMC- 5 YEARS POST WARRANTY (i.e. 6th to 10th Years) to be submitted by OEM on Letterhead

Under CMC:- Repair, Maintenance & Replacement of Spares without any additional cost

The CMC for 5 years i.e. from 6th to 10th year shall cover CMC on all items supplied by third party. Including upgradation of software (If any)

Details of Existing equipment available in the department of Buy - Back Purpose

1) Make:- Olympus

2) Model:-GIF-V, GIF-Q 160, CF Q 160L, JF-V, PVM-14N6E, CV-160, PSD-20 & KV-4

3) Sr. No.:- 2922161, 2102059, 2101343, 2910194,6004337,7000998,7000914, CE197 & 990775
2

4) Year of Purchase:- 2000-2001

Kindly comply all the Technical Specifications parameters as mentioned in buyer Added bid Specific ATC also.

Lower GI Scope as per Annexure: "A"

Lower GI video scope with HD image quality: Should have following specifications:

1. High Definition image quality
2. Optical Chromoendoscopy such as NBI/BLI/i-Scan OE
3. Scope should be fully immersible in disinfectant solution.
4. Additional/auxiliary channel for water jet function
5. Pre-freezer function for sharper images
6. Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump or separate leakage testing device.
7. Scope ID function
8. Four or more remote control switches on control body
9. Water proof connector which is fully immersible in disinfectant solution
10. Better scope insertion & control features such as Passive Bending, High Force Transmission, Variable Stiffness or equivalent
11. Additional/auxiliary channel for water jet function
12. Field of view- 170° or more
13. Direction of view -Forward view
14. Depth of field - 2-100mm or better
15. Distal end outer diameter of 13.2 mm or less
16. Insertion tube outer diameter of 12.8 mm or less
17. Working length of 1680 mm or more
18. Channel inner diameter of 3.7 mm or more
19. Minimum visible distance of 4.0 mm or better from the distal end
20. Angulation range: Up -180° , Down -180° , Right-160° & Left-160° or more
21. Total length of 2005 mm or more

Annexure:"B'

Duodenovideoscope: Should have following specifications:

1. Lighter and possess high resolution image quality
2. Unique dual locking mechanism: Center lock and side lock
3. Optical Chromoendoscopy such as NBI/BLI/i-Scan OE
4. Scope ID function
5. Four or more remote control switches on control body
6. Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump.
7. Field of view of 100° or more
8. Direction of view- Backward side viewing 5° or better
9. Depth of field of 5-60 mm or better
10. Distal end outer diameter of 13.7 mm or less
11. Insertion tube outer diameter of 11.3 mm or less
12. Working length of 1240 mm or more
13. Instrument channel inner diameter of 4.2 mm or more
14. Minimum visible distance of 10 mm or better

15. Angulation range - Up: 120, Down : 90° Right: 110° , Left : 90° or better

16. Total length of 1550 mm or more

2.20 Buyer Added Bid Specific ATC:

Buyer uploaded ATC document [Click here to view the file.](#)

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.