

1	ईमेल आईडी Email ID : cdmo-gh-lwd@gujarat.gov.in संपर्क Contact : 2674-255008- जीएसटीआईएन GSTIN : - पता Address : Janeral Hospital Lunawada Cottage Chowkdi Lunawada Mahisagar, PANCH MAHALS, GUJARAT-389230, India	chempack Single sided Sterilization Indicator Tapes For Medical Purpose 55 meter	-	3	14-Mar-2024	29-Mar-2024
Product Specification for chempack Single sided Sterilization Indicator Tapes For Medical Purpose 55 meter						
विनिर्देश Specification	उप-विनिर्देश Sub-Spec	मूल्य Value				
PRODUCT TECHNICAL FEATURES	Method of sterilization	ETO				
	Colour of tape before steriliation	Beige				
	Colour of tape after steriliation/autoclave	Brown				
	Change of the integrated chemical indicator on the tape	Its colour				
	After the sterilization process is complete, the tape can easily be removed with no adhesive residue	Yes				
	Tape provides an indication that the pack has passed through a sterilization cycle	Yes				
	Storage temperature range (Deg C)	Any applicable text				
	Storage humidity range (%RH)	Any applicable text				
	Sterilization temperature (Deg C)	Any applicable text				
	Sterilization relative humidity (RH) (%)	Any applicable text				
	Sterilization time (Minutes)	Any applicable text				
	Exposition	Any applicable text				
	Shelf life of tape (year)	2				
GENERAL FEATURES	Product	Sterilization/autoclave indicator tapes for Medical purposes				
	Good resistant to heat and moisture	Yes				
	Tape shall be Lead free and Non - toxic	Yes				
	Strong tape, firmly holds to package before, during and after sterilization cycle	Yes				
	Tape shall stick well on paper, board, metal, plastic, glass etc	Yes				
	Type	Single sided				
	Usage	Sealing or packaging to distinguish between sterilized and non-sterilized units during storage				
	Objective	To secure the packs wrapped with woven, non-woven, treated woven, paper and paper/plastic wraps that go into sterilizer and verify the packs that have been exposed to Hydrogen Peroxide. Indicator color does not fade when exposed to light.				
	Easily removable, residue free	Yes				
Highly resistant to break or tear	Yes					
DIMENSIONS, MATERIAL & WEIGHT	Tape length (mtrs)	55 meter				
	Tape width (mm)	18 millimeter				
	Core diameter of roll (mm)	Any applicable text				
	Material of tape backing	Semi-creaped paper				
	Other side of tape	Chemical indicator Stripes or words that change in colour on sterilization				
	Weight (gm)	100 gram				
STANDARD	Conformity to Standard	ISO 18472:2006 Sterilization of Health care products- Biological and Chemical indicators				
	ISI Marked	No				
	CM/L Number (if product is ISI marked)	0				
PACKING	Packing	Individually packed in peel-to-open paper/polythene pouch				
CERTIFICATION & REPORTS	Submission of test report on Form 39 or from Central Government /NABL / ILAC accredited Lab covering all the parameters declared	Yes				
	Certification available	ISO				
	Product conformity certificate is to be provided to the buyer at the time of supply	Yes				

टिप्पणी | Note:: Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.

ईपीबीजी विवरण | ePBG Detail

NA

नियम और शर्तें | Terms and Conditions

1. General Terms and Conditions-

1.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.

1.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.