

1	ईमेल आईडी Email ID : buycon20.hfwdhp.hp@gembuyer.in संपर्क Contact : 01972-222999- जीएसटीआईएन GSTIN : - पता Address : Medical college Hamirpur, HAMIRPUR (HP), HIMACHAL PRADESH-177001, India	Bio-AGE Type1 and Type 2 High quality water purification system	-	1	18-Mar-2024	02-Apr-2024
Product Specification for Bio-AGE Type1 and Type 2 High quality water purification system						
विनिर्देश Specification		उप-विनिर्देश Sub-Spec			मूल्य Value	
Standards	Certification for product		CE			
	Certification number and date		4248 DATED 18/04/2016			
	Quality system certifications of manufacturer		ISO 13485 and 9001			
	Confirmity to standard for for electrical safety latest amendment		IEC- 60601 or BIS Equivalent			
Performance Parameters	Purpose		Lab applications for HPLC/Ion Chromatography / LCMS/Titration			
	Prefiltration RO Unit supplied		Yes			
	In put quality in TDS for prefiltration unit		To be less than 2000 TDS			
	Out put quality from prefiltration RO Unit in TDS		300<			
	Out put flow rate for pre filtration unit		Minimum 50 litre per hour to be provided			
	In case pre filtration unit provided capacity of water storage tank from pre filtration unuttb		Minimum 50 litre			
	In put water for main RO unit		Less than 300 TDS			
	Out put water from main RO Unit		Less than 5 TDS			
	Output flow rate from main RO unit		Minimum 25 litre per hour			
	Water tank capacity of main RO UNIT		50 Litre minimum			
	Technology of water purification used to achieve ultra purified water		Ion exchange resin technology			
	Type of water produced by system		Type1 and Type 2			
	Resistivity of water out put		10-15 mega ohms .cms or better @25 degree celsius (For Type I)			
	Toc (ppb)		<30 (Type 2)			
	Particles after filtering		no particles with size greater than 0.2 micro meter			
	Endotoxins		<0.001 EU per ml			
	Bacteria		<0.1 CFU per ml			
	RNAse		<10 picogram per ml			
	DNAse		<5 picogram per ml			
	Flow rate in litres per hour		Minimum 03 liters per hr			
Display provided		LED				
Display parameters		Resistivity and level of water in reservoir				
Active carbon beads Electro Deionization Module (EDI)		Yes				
Provision of UV in storage tank		Yes				
Miscellaneous Parameters	Warranty in years		5			
	For easy understanding of main operations Built in quick reference guide in system		Yes			
	Installation and demonstration to be given before acceptance		Yes			
	Instruction manual to be supplied		Yes			
	Number of installations of the Machine machines in Central /State/PSU Govt Hospitals(Note: Seller should supply a performace certificate of the device to the buyer if demanded after placement of order)		More than 3			
	OEM/Reseller (if supplied to reseller shall ensure uninterputed availability of all spares for 10 years)		Yes			
	Availability of toll free facility for technical support maintained by OEM or authorized agencies		Yes			
	User/Technical/Maintenance manuals in original to be supplied in English in hard and soft copy		Yes			
	Demonstration of equipment and training to be provided after completing supplies before acceptance		Yes			
	The Principal Manufacturer must have direct Presence/approved service center In India		Yes			
QA, QC and other tests as per NABH / NABL shall be carried out by OEM or authorized agency of OEM		Yes				

ईपीबीजी विवरण | ePBG Detail

NA

नियम और शर्तें | Terms and Conditions

1. General Terms and Conditions-

1.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.

1.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.