

1	संपर्क Contact : 0477-2280300- जीएसटीआईएन GSTIN : 27AAEAT4818Q3ZQ पता Address : E-BLOCK, 2ND FLOOR, GOVT.T.D.M.C HOSPITAL COMPLEX, VANDANAM, ALAPPUZHA, KERALA-688005, ALAPPUZHA, KERALA-688005, India	LABWAN Brushless DC Motor Vortex Mixers	-	1	16-Apr-2024	01-May-2024
---	--	--	---	---	-------------	-------------

Product Specification for LABWAN Brushless DC Motor Vortex Mixers

विनिर्देश Specification	उप-विनिर्देश Sub-Spec	मूल्य Value
WARRANTY	Warranty on machine	2 year
DISCREPTION	Type of Drive Motor	Brushless DC Motor
	Display Option/Output Read out Option	Analog
	Material of Mixing Head/vortex cup	Rubber
	Type of Vortex Mixer	Mini Vortex Mixer
	Type of Controller	Microprocessor Controlled
	Timer Setting (Increment step)	0.0001 minute
	Base material	Die cast aluminium
	Number of Sample tubes testable simultaneously	Multi-Tube
	Weight (kg)	2 kilogram
	User Interface	Analog Front Panel
	Largest Volume of the tube testable	200 milliliter
	Finish of Body material	Powder coated
	Body Material	Plastic
	Noise Level (maximum) (db)	20 decibel
Maximum Load capacity	0.1 kilogram	
FEATURES	Speed increment step,if variable speed (RPM)	3000
	IP rating or Ingress protection rating	21
	Activation mode	Press Activated and Continuous run
	Speed mode	Variable Speed
	Type of Data Logging option	USB
	Number of Programmable/Programmed modes	0
	Maximum Speed (RPM)	3000
	Minimum Speed (RPM)	1
	Timer Provided	Yes
	Input Current (Amp)	0.01
	Pulse mode	No
	In built Counter balance	Yes
	Programmable Pulse mode	No
	Speed control Accuracy (+/-) (RPM)	10
	Suction Cup in the base	Yes
	Maximum Diameter of the tube testable (mm)	75
Accessories and spares included in standard pack	Conical/Volumetric	
Auto Cutoff	Yes	
DIMENSIONS	Orbital Diameter	2 millimeter
	External Dimension: Height	17 centimeter
	External Dimension: Width	13 centimeter
	External Dimension: Depth	19 centimeter
POWER	Power Consumption (W)	300 Watt
	Input Voltage (V) with number of Phases	230
	availability of test report (to be furnished to Buyer by Seller on demand of the former) from NABL/ILAC/Govt laboratory	No
	Test Report Number and date	NA

CERTIFICATIONS	Name and address of the Lab issuing test report	NA
	Product CE certified	Yes
	Firm certified to ISO9001:2008 or latest	YES
	If CE certified, EN/IEC directives	NA
OPERATING PARAMETERS	Ambient operating Temperature (0 to 50 Degree Celsius)	YES
<p>टिप्पणी Note:: Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.</p>		
<p>ईपीबीजी विवरण ePBG Detail</p>		
<p>NA</p>		
<p>नियम और शर्तें Terms and Conditions</p>		
<p>1. General Terms and Conditions-</p> <p>1.1 This contract is governed by the General Terms and Conditions, conditions stipulated to this Product/Service as provided in the Marketplace.</p> <p>1.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable</p>		
<p>नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।</p> <p>Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.</p>		